



## Management-Handbuch Produktion



Secure the Handler with air pad foil and safe the carrousel with the transport lock and etafon pads.



Put air pad foil between the glass- and Frontdoor.

Packing\_STX1000.doc 2010.10.18 Page 1 of 3



## Management-Handbuch Produktion


Pack the Unit in protection foil.





Screw the "Etafon" on the palette and then put the unit on.


Attention: There has to be 20mm space between the Incubator and the palette.





## Management-Handbuch Produktion


The screwing: Screw each side wall together with the palette with two screw at the outer bottom corner at first and then with one screw at the middle.





Next step, screw the sidewalls at the upper corners together, also with two screws. Place screws the whole edge with a distance of 350 mm. The lid has to be screwed once together with each sidewall and each corner. Take counter sunk head wood screws 4.5x50 with non-stop thread. Screw in the screw heads flush. Don't counterbore the screws in the wood!





Tape six corner pads on the front wall and screw it together with the palette.

Screw the sidewalls with the palette together. Push the incubator completely to the front, it has to be close to the corner pads.

Wind the transfer station in air pad foil and put this between the incubator and the ground.

Tape the back corner pads on the wood cube and screw together with the sidewall. The unit has to be close between the front and the back corner pads so it can't move during the transport.



Screw the back wall and the lid.

