



Title:	Removal STX40 Lift Column			ID:
				0013
Date in:	Response:	Model:	Author:	
2007-02-14	2007-02-14	STX40	CMa	

STX40 Removal Lift Column

1. Move lift wagon up and secure



2. Mark Lift column position



3. Remove lift cable from cable clip located in the upper area of the left sidewall inside the incubation chamber.

4. Loosen the four (4) screws that hold the lift column and remove lift column

